

**SPESIFIKASI TEKNIS**  
**ELITECH MEDICAL UV STERILIZER GET 338-U**

**Merk** : ELITECH  
**Nomor Izin Edar** : KEMENKES RI AKD 20903800291

PRODUSEN : PT. SINKO PRIMA ALLOY  
 SERTIFIKAT PRODUKSI : HK.02.06/VI/119/2016  
 ALAMAT : JL. Tambak Osowilangun No.61  
 Pergudangan Osowilangun Permai Blok E7-E8, Surabaya

<b>SPESIFIKASI TEKNIS</b>		
Model		GET 338 U
Power Supply Input		AC 220V,50Hz
Daya Terpakai		± 45W
Kapasitas Total Ruangan		38 L
Daya Tampung Rak Atas & Rak Bawah		3 kg dan 3 kg
Dimensi		435mm(L) x 375mm(W) x 450mm(H)
Berat		9 Kg
Ultra Violet	Intensitas	≥60 μW/cm <sup>2</sup>
	Waktu Sterilisasi	60 menit
	Hasil Uji Sterilisasi	>90% (Steril)
Auto UV I dan UV II	Intensitas	≥60 μW/cm <sup>2</sup> (UV)
	Waktu Sterilisasi	120 menit
	Hasil Uji Sterilisasi	>90% (Steril)
Normal Operating	Ambient Temperature	+5 °C - +35 °C
	Relative humidity	≤ 80%
	Athmosphere	86 kPa - 106 kPa
Transport Operating	Ambient Temperature	-10 °C - +55 °C
	Relative humidity	≤ 95%
	Athmosphere	50 kPa - 106 kPa